



Product Service

Attestation of Conformity

No. E8A 101747 0003 Rev. 00

Holder of Certificate:	AR CO., LTD. 709, 3DA, 21, Gongdan 1-daero 260beonan-gil Siheung-si, Gyeonggi-do 15104 REPUBLIC OF KOREA	
Name of Object:	Electrical equ. for measurement, control and laboratory use (Negative Pressure Air Unit)	
Model(s):	ANPU-15	
Description of Object:	Rated input voltage:	220-240 V a.c.
	Rated input frequency:	50/60 Hz
	Rated input current:	1.5 A
Tested according to:	EN 61000-6-4:2007/A1:2011 EN 61000-6-2:2005 EN 61000-3-2:2014 EN 61000-3-3:2013	

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: DST-20E-0524

Date, 2020-07-10

(Byung-Soo Kang)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.